

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.1uS / 0.0ppm  
CYCLE NUMBER 005635  
9:04 31/05/2023  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 1:58  
PRESSURIZING 2:01  
STERILIZING 3:09  
136.5 C 222kPa 3:09  
136.2 C 221kPa 3:39  
136.4 C 223kPa 4:09  
136.4 C 221kPa 4:39  
136.4 C 223kPa 5:09  
136.4 C 220kPa 5:39  
136.4 C 224kPa 6:09  
136.4 C 224kPa 6:39  
136.4 C 224kPa 6:39  
VENTING 6:39  
AIR DRYING 7:28  
CYCLE COMPLETE 10:44

-----  
Digital Signature #  
B372833DD8FE8224  
-----