
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001839
12:56 12/07/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:12
PRESSURIZING 2:19
STERILIZING 3:25
136.4 C 226kPa 3:25
136.2 C 224kPa 3:55
136.4 C 226kPa 4:25
136.4 C 226kPa 4:55
136.4 C 222kPa 5:25
136.2 C 223kPa 5:55
136.3 C 225kPa 6:25
136.4 C 224kPa 6:55
136.4 C 224kPa 6:55
VENTING 6:55
AIR DRYING 7:44
CYCLE COMPLETE 7:52

Digital Signature #
8E19E3D9C4257A87
