
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000110
9:40 03/09/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:16
PRESSURIZING 4:17
STERILIZING 5:54
136.4 C 224kPa 5:54
136.2 C 224kPa 6:24
136.3 C 226kPa 6:54
136.1 C 223kPa 7:24
136.3 C 224kPa 7:54
136.2 C 224kPa 8:24
136.4 C 226kPa 8:54
136.4 C 226kPa 9:24
136.4 C 226kPa 9:24
VENTING 9:24
AIR DRYING 10:28
CYCLE COMPLETE 25:28

Digital Signature #
229030CA71A910A8
