
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.5uS / 0.3ppm
CYCLE NUMBER 000976
14:42 09/03/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:06
PRESSURIZING 2:13
STERILIZING 3:14
136.0 C 229kPa 3:14
136.4 C 232kPa 3:45
136.4 C 233kPa 4:15
136.4 C 233kPa 4:45
136.4 C 233kPa 5:15
136.4 C 233kPa 5:45
136.3 C 232kPa 6:15
136.3 C 233kPa 6:45
136.3 C 233kPa 6:45
VENTING 6:45
AIR DRYING 7:36
CYCLE COMPLETE 7:38

Digital Signature #
908C7034929159E3
