
Midmark M11 - v1.0.0
Total Cycles: 1379

Name: RITTER 1
Sterilizer ID: V2178198
11/17/2021 03:07 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 178.7 F 0.0
2:00 178.4 F 3.5
4:00 191.2 F 7.7
6:00 214.1 F 12.7
8:00 222.7 F 7.1
10:00 236.5 F 12.4
12:00 237.1 F 9.6
14:00 237.4 F 8.8
16:00 249.6 F 15.0
18:00 261.8 F 21.9
20:00 272.7 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.2
1:00 272.9 F 29.0
2:00 271.9 F 28.3
3:00 272.1 F 28.8
4:00 272.7 F 28.9

Min 271.5 F 28.1
Max 273.0 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.3 F 0.0
5:00 223.5 F 0.0
10:00 235.3 F 0.0
15:00 232.7 F 0.0
20:00 226.7 F 0.0
25:00 225.3 F 0.0
30:00 225.0 F 0.0
Drying Complete

FILLING: 1:25
HEATING: 20:20
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 00:58:54
CYCLE COMPLETE

11/17/2021 04:05 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____