

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3937

Name: RITTER 1  
Sterilizer ID: V2178198  
10/23/2023 01:39 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 140.2 F 0.0  
2:00 156.2 F 1.5  
4:00 158.3 F 4.5  
6:00 190.2 F 8.1  
8:00 213.7 F 12.3  
10:00 218.9 F 5.0  
12:00 229.5 F 9.1  
14:00 239.9 F 13.5  
16:00 237.8 F 10.1  
18:00 237.5 F 8.9  
20:00 247.4 F 13.8  
22:00 258.3 F 19.7  
24:00 268.6 F 26.2

STERILIZING  
mm:ss Degrees PSI  
0:00 272.0 F 28.6  
1:00 272.7 F 28.9  
2:00 272.7 F 29.1  
3:00 271.3 F 28.7  
4:00 272.1 F 28.7

Min 271.2 F 28.5  
Max 272.9 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 187.5 F 0.0  
5:00 238.7 F 0.0  
10:00 223.4 F 0.0  
15:00 228.0 F 0.0  
20:00 219.4 F 0.0  
25:00 220.0 F 0.0  
30:00 232.4 F 0.0  
Drying Complete

FILLING: 1:14  
HEATING: 25:48  
STERILIZING: 4:00  
VENTING: 3:40  
DRYING: 30:00  
TOTAL CYCLE: 01:04:42  
CYCLE COMPLETE

10/23/2023 02:43 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_