

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1233

Name: RITTER 1  
Sterilizer ID: V2178198  
10/06/2021 06:33 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 72.2 F 0.0  
2:00 75.5 F 0.0  
4:00 87.6 F 0.0  
6:00 127.3 F 0.0  
8:00 157.2 F 0.1  
10:00 183.8 F 0.2  
12:00 202.3 F 0.3  
14:00 215.3 F 3.8  
16:00 223.8 F 6.5  
18:00 232.2 F 9.8  
20:00 238.3 F 11.0  
22:00 240.0 F 10.9  
24:00 236.5 F 8.7  
26:00 241.1 F 10.1  
28:00 249.2 F 14.4  
30:00 261.0 F 21.0  
32:00 271.8 F 28.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.9 F 27.8  
1:00 272.3 F 28.4  
2:00 272.2 F 28.1  
3:00 271.8 F 28.1  
4:00 272.7 F 28.5  
  
Min 271.5 F 27.7  
Max 272.8 F 28.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 184.2 F 0.0  
5:00 207.7 F 0.0  
10:00 206.2 F 0.0  
15:00 208.6 F 0.0  
20:00 206.1 F 0.0  
25:00 202.1 F 0.0  
30:00 207.9 F 0.0  
  
Drying Complete

FILLING: 0:55  
HEATING: 32:27  
STERILIZING: 4:00  
VENTING: 3:18  
DRYING: 30:00  
TOTAL CYCLE: 01:10:39  
CYCLE COMPLETE

10/06/2021 07:45 AM  
ACCEPTED: X  
REJECTED:  
Operator: U01