

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3061

Name: RITTER2  
Sterilizer ID: V2313228  
10/30/2023 08:21 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 68.3 F 0.0  
2:00 78.8 F 0.0  
4:00 89.6 F 0.3  
6:00 125.8 F 1.5  
8:00 140.8 F 4.4  
10:00 170.2 F 7.9  
12:00 203.8 F 11.8  
14:00 215.9 F 4.1  
16:00 226.3 F 7.8  
18:00 235.8 F 11.9  
20:00 233.3 F 8.3  
22:00 242.3 F 12.4  
24:00 243.1 F 12.2  
26:00 255.3 F 18.5  
28:00 266.0 F 25.2  
30:00 271.7 F 29.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 29.4  
1:00 272.3 F 29.3  
2:00 272.5 F 29.6  
3:00 272.8 F 29.9  
4:00 272.1 F 29.9

Min 271.5 F 29.3  
Max 272.9 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 186.8 F 0.0  
5:00 232.7 F 0.0  
10:00 228.0 F 0.0  
15:00 230.1 F 0.0  
20:00 223.2 F 0.0  
25:00 228.5 F 0.0  
30:00 226.9 F 0.0  
Drying Complete

FILLING: 0:49  
HEATING: 30:29  
STERILIZING: 4:00  
VENTING: 3:14  
DRYING: 30:00  
TOTAL CYCLE: 01:08:32  
CYCLE COMPLETE

10/30/2023 09:29 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_