
Midmark M11 - v1.0.5
Total Cycles: 2862

Name: RITTER2
Sterilizer ID: V2313228
09/11/2023 02:16 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 174.3 F 0.0
2:00 172.7 F 3.2
4:00 185.4 F 7.3
6:00 213.8 F 11.8
8:00 221.4 F 6.7
10:00 230.1 F 11.8
12:00 234.9 F 9.2
14:00 238.4 F 10.2
16:00 247.2 F 14.2
18:00 258.6 F 20.5
20:00 269.1 F 27.2

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.9
1:00 272.5 F 29.6
2:00 272.9 F 29.8
3:00 271.7 F 29.3
4:00 271.9 F 29.8

Min 271.0 F 28.8
Max 273.2 F 30.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 197.4 F 0.0
5:00 216.1 F 0.0
10:00 209.2 F 0.0
15:00 203.9 F 0.0
20:00 201.9 F 0.0
25:00 201.6 F 0.0
30:00 203.0 F 0.0
Drying Complete

FILLING: 1:44
HEATING: 21:09
STERILIZING: 4:00
VENTING: 3:40
DRYING: 30:00
TOTAL CYCLE: 01:00:33
CYCLE COMPLETE

09/11/2023 03:16 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____