
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 007172
9:24 09/04/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:55
PRESSURIZING 3:56
STERILIZING 4:38
136.0 C 220kPa 4:38
136.1 C 221kPa 5:08
135.9 C 221kPa 5:38
136.4 C 222kPa 6:08
136.1 C 219kPa 6:38
136.3 C 220kPa 7:08
136.3 C 222kPa 7:38
136.4 C 223kPa 8:08
136.4 C 223kPa 8:08
VENTING 8:09
AIR DRYING 8:59
CYCLE COMPLETE 9:01

Digital Signature #
E98AA5257D29E0F9
