
Midmark M11 - v1.0.0
Total Cycles: 1683

Name: RITTER 1
Sterilizer ID: V2178198
03/01/2022 08:13 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 75.3 F 0.0
2:00 85.5 F 0.0
4:00 89.8 F 0.5
6:00 128.6 F 1.9
8:00 158.5 F 3.7
10:00 181.4 F 6.8
12:00 212.1 F 10.6
14:00 218.7 F 5.5
16:00 226.5 F 7.0
18:00 237.0 F 11.1
20:00 230.9 F 6.7
22:00 244.5 F 12.8
24:00 241.0 F 10.7
26:00 254.2 F 17.2
28:00 265.3 F 23.9

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.1
1:00 272.6 F 28.7
2:00 271.2 F 28.0
3:00 272.6 F 28.7
4:00 271.3 F 27.9

Min 271.1 F 27.9
Max 272.7 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 189.1 F 0.0
5:00 207.1 F 0.0
10:00 208.3 F 0.0
15:00 207.9 F 0.0
20:00 186.8 F 0.0
25:00 196.6 F 0.0
30:00 168.1 F 0.0
Drying Complete

FILLING: 0:52
HEATING: 29:48
STERILIZING: 4:00
VENTING: 2:52
DRYING: 30:00
TOTAL CYCLE: 01:07:33
CYCLE COMPLETE

03/01/2022 09:49 AM
ACCEPTED: X
REJECTED:
Operator: U02