

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1310

Name: RITTER 1  
Sterilizer ID: V2178198  
10/27/2021 02:24 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 136.6 F 0.0  
2:00 138.6 F 1.0  
4:00 163.2 F 3.1  
6:00 181.5 F 7.0  
8:00 203.0 F 11.8  
10:00 216.8 F 5.1  
12:00 227.1 F 9.7  
14:00 234.8 F 9.5  
16:00 240.8 F 11.7  
18:00 238.7 F 9.9  
20:00 251.8 F 16.3  
22:00 263.1 F 22.9  
24:00 272.7 F 29.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.7 F 28.6  
1:00 272.8 F 29.3  
2:00 271.6 F 28.5  
3:00 272.7 F 29.4  
4:00 272.1 F 28.8

Min 271.6 F 28.4  
Max 272.9 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 183.2 F 0.0  
5:00 209.5 F 0.0  
10:00 214.5 F 0.0  
15:00 214.8 F 0.0  
20:00 218.3 F 0.0  
25:00 218.5 F 0.0  
30:00 216.5 F 0.0  
Drying Complete

FILLING: 1:22  
HEATING: 24:17  
STERILIZING: 4:00  
VENTING: 3:22  
DRYING: 30:00  
TOTAL CYCLE: 01:03:01  
CYCLE COMPLETE

10/27/2021 03:26 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_