

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3396

Name: RITTER 1  
Sterilizer ID: V2178198  
06/15/2023 11:56 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 196.0 F 0.0  
2:00 192.0 F 1.5  
4:00 216.2 F 7.1  
6:00 225.6 F 12.6  
8:00 230.6 F 8.7  
10:00 239.0 F 10.5  
12:00 243.8 F 12.6  
14:00 242.5 F 11.7  
16:00 256.0 F 18.5  
18:00 267.4 F 25.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 28.3  
1:00 272.5 F 29.0  
2:00 271.3 F 28.3  
3:00 272.3 F 28.9  
4:00 271.2 F 28.2  
  
Min 271.2 F 28.2  
Max 272.6 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 203.2 F 0.0  
5:00 221.7 F 0.0  
10:00 232.3 F 0.0  
15:00 238.0 F 0.0  
20:00 234.5 F 0.0  
25:00 226.3 F 0.0  
30:00 222.5 F 0.0  
Drying Complete

FILLING: 1:56  
HEATING: 19:15  
STERILIZING: 4:00  
VENTING: 3:24  
DRYING: 30:00  
TOTAL CYCLE: 00:58:35  
CYCLE COMPLETE

06/15/2023 12:54 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_