
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 001621
17:48 20/11/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:02
PRESSURIZING 3:03
STERILIZING 4:03
136.4 C 225kPa 4:03
136.4 C 227kPa 4:33
136.2 C 225kPa 5:03
136.4 C 227kPa 5:33
136.4 C 227kPa 6:03
136.2 C 224kPa 6:33
136.4 C 227kPa 7:03
136.4 C 226kPa 7:33
136.4 C 226kPa 7:33
VENTING 7:33
AIR DRYING 8:28
CYCLE COMPLETE 22:58

Digital Signature #
AC49803AF557FCA1
