

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 000433  
8:37 28/01/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:34  
PRESSURIZING 2:43  
STERILIZING 3:46  
136.4 C 223kPa 3:46  
136.4 C 233kPa 4:16  
136.4 C 229kPa 4:46  
136.2 C 230kPa 5:16  
136.2 C 230kPa 5:46  
136.4 C 231kPa 6:16  
136.1 C 230kPa 6:46  
136.4 C 231kPa 7:16  
136.4 C 231kPa 7:16  
VENTING 7:16  
AIR DRYING 8:09  
CYCLE COMPLETE 8:14

-----  
Digital Signature #  
1EB0603BE06E2D8A  
-----