
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 005997
9:16 18/07/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:20
PRESSURIZING 2:24
STERILIZING 3:16
136.0 C 220kPa 3:16
136.2 C 223kPa 3:46
136.4 C 224kPa 4:16
136.1 C 219kPa 4:46
136.4 C 224kPa 5:16
136.0 C 219kPa 5:46
136.4 C 224kPa 6:16
136.4 C 225kPa 6:46
136.4 C 225kPa 6:46
VENTING 6:46
AIR DRYING 7:37
CASSETTE REMOVED
CYCLE COMPLETE 16:12

Digital Signature #
935DA7147B835F74
