
Midmark M11 - v1.0.5
Total Cycles: 2204

Name: RITTER2
Sterilizer ID: V2313228
03/31/2023 09:22 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 214.3 F 0.0
2:00 209.9 F 4.8
4:00 224.1 F 11.8
6:00 236.1 F 10.4
8:00 238.7 F 10.2
10:00 240.0 F 10.5
12:00 257.0 F 19.1

C103: HEATUP MODE
STOP PRESSED
CHAMBER TEMP 256F
CHAMBER PRESS 16.9PSI
** ITEMS NOT STERILE **

03/31/2023 09:39 AM