
Midmark M11 - v1.0.0
Total Cycles: 3041

Name: RITTER 1
Sterilizer ID: V2178198
03/27/2023 04:27 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 174.7 F 0.0
2:00 172.8 F 4.2
4:00 190.4 F 8.6
6:00 208.7 F 11.9
8:00 221.3 F 7.4
10:00 235.3 F 12.5
12:00 235.4 F 9.5
14:00 237.5 F 8.9
16:00 248.1 F 14.5
18:00 260.6 F 21.4
20:00 271.9 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.5
1:00 272.4 F 29.0
2:00 271.5 F 28.6
3:00 272.4 F 29.1
4:00 271.5 F 28.4

Min 271.4 F 28.4
Max 272.5 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.2 F 0.0
5:00 209.2 F 0.0
10:00 211.6 F 0.0
15:00 215.1 F 0.0
20:00 228.1 F 0.0
25:00 231.6 F 0.0
30:00 227.3 F 0.0
Drying Complete

FILLING: 1:08
HEATING: 20:33
STERILIZING: 4:00
VENTING: 3:18
DRYING: 30:00
TOTAL CYCLE: 00:58:59
CYCLE COMPLETE

03/27/2023 05:25 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____