
Midmark M11 - v1.0.0
Total Cycles: 3404

Name: RITTER 1
Sterilizer ID: V2178198
06/19/2023 10:56 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 195.0 F 0.0
2:00 188.8 F 2.2
4:00 202.2 F 6.9
6:00 219.4 F 12.4
8:00 227.7 F 7.8
10:00 240.4 F 13.1
12:00 239.8 F 10.8
14:00 237.1 F 9.0
16:00 251.4 F 15.9
18:00 262.9 F 22.5
20:00 272.4 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.2
1:00 272.5 F 28.7
2:00 272.1 F 28.7
3:00 272.1 F 28.4
4:00 272.4 F 29.0

Min 271.5 F 28.1
Max 272.9 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.7 F 0.0
5:00 212.3 F 0.0
10:00 222.7 F 0.0
15:00 212.7 F 0.0
20:00 209.6 F 0.0
25:00 209.4 F 0.0
30:00 209.2 F 0.0
Drying Complete

FILLING: 1:54
HEATING: 20:13
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:59:17
CYCLE COMPLETE

06/19/2023 11:55 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____