
Brampton, ON
Canada

STATIM 5000 S501R713
SN 511019D00004
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001416
10:30 09/09/2021
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:02
PRESSURIZING 5:28
STERILIZING 6:50
136.0 C 220kPa 6:50
136.5 C 224kPa 7:20
136.4 C 224kPa 7:50
136.5 C 221kPa 8:20
136.3 C 221kPa 8:50
136.4 C 223kPa 9:21
136.4 C 224kPa 9:51
136.2 C 222kPa 10:20
VENTING 10:20
AIR DRYING 11:20
CYCLE COMPLETE 11:28

Digital Signature #
5519E218C72C5423
