
Midmark M11 - v1.0.0
Total Cycles: 1314

Name: RITTER 1
Sterilizer ID: V2178198
10/28/2021 10:20 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 178.7 F 0.0
2:00 174.4 F 3.1
4:00 196.9 F 7.2
6:00 216.8 F 12.7
8:00 226.6 F 7.4
10:00 239.9 F 12.9
12:00 238.7 F 10.3
14:00 235.7 F 8.5
16:00 251.2 F 15.8
18:00 263.6 F 22.9
20:00 271.3 F 28.1

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.1
1:00 272.4 F 28.8
2:00 271.2 F 28.1
3:00 272.7 F 28.9
4:00 271.6 F 28.2

Min 271.2 F 28.0
Max 272.7 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 181.2 F 0.0
5:00 229.4 F 0.0
10:00 221.7 F 0.0
15:00 218.3 F 0.0
20:00 217.7 F 0.0
25:00 221.6 F 0.0
30:00 225.3 F 0.0
Drying Complete

FILLING: 1:50
HEATING: 20:01
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 00:59:00
CYCLE COMPLETE

10/28/2021 11:28 AM
ACCEPTED: X
REJECTED:
Operator: U04