
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001409
12:11 10/02/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:57
PRESSURIZING 3:02
STERILIZING 4:17
136.0 C 220kPa 4:17
136.4 C 225kPa 4:47
136.4 C 220kPa 5:17
136.0 C 222kPa 5:47
136.4 C 224kPa 6:17
136.1 C 223kPa 6:47
136.4 C 224kPa 7:17
136.3 C 223kPa 7:47
136.3 C 223kPa 7:47
VENTING 7:47
AIR DRYING 8:38
CYCLE COMPLETE 23:38

Digital Signature #
058B2592C22F94CE
