
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001293
9:24 13/10/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:56
PRESSURIZING 3:59
STERILIZING 5:36
136.0 C 220kPa 5:36
136.3 C 223kPa 6:06
136.2 C 222kPa 6:36
136.4 C 225kPa 7:06
136.4 C 222kPa 7:36
136.2 C 223kPa 8:06
136.4 C 224kPa 8:36
136.4 C 224kPa 9:06
136.4 C 224kPa 9:06
VENTING 9:06
AIR DRYING 10:06
CYCLE COMPLETE 25:06

Digital Signature #
AB9BD758ED31987E
