

-----  
Midmark M11 - v1.0.5  
Total Cycles: 4312

Name: RITTER2  
Sterilizer ID: V2313228  
01/29/2025 02:42 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 155.0 F 0.0  
2:00 156.9 F 2.8  
4:00 168.7 F 6.7  
6:00 191.1 F 11.4  
8:00 212.9 F 5.3  
10:00 220.9 F 10.1  
12:00 230.3 F 8.6  
14:00 238.4 F 12.4  
16:00 239.1 F 11.2  
18:00 251.7 F 17.4  
20:00 262.5 F 23.8  
22:00 272.5 F 30.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 29.8  
1:00 271.1 F 29.7  
2:00 271.2 F 30.1  
3:00 272.5 F 31.1  
4:00 272.9 F 31.1

Min 271.0 F 29.6  
Max 273.2 F 31.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 196.0 F 0.0  
5:00 194.7 F 0.0  
10:00 210.7 F 0.0  
15:00 209.2 F 0.0  
20:00 205.9 F 0.0  
25:00 204.5 F 0.0  
30:00 202.0 F 0.0  
Drying Complete

FILLING: 1:20  
HEATING: 22:26  
STERILIZING: 4:00  
VENTING: 3:55  
DRYING: 30:00  
TOTAL CYCLE: 01:01:42  
CYCLE COMPLETE

01/29/2025 03:43 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_