
Midmark M11 - v1.0.0
Total Cycles: 3101

Name: RITTER 1
Sterilizer ID: V2178198
04/12/2023 12:56 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 202.0 F 0.0
2:00 193.9 F 4.2
4:00 213.8 F 9.5
6:00 221.8 F 6.4
8:00 236.1 F 10.8
10:00 234.7 F 8.4
12:00 240.3 F 10.1
14:00 248.2 F 14.3
16:00 261.0 F 21.4
18:00 272.5 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.3
1:00 272.6 F 29.0
2:00 271.5 F 28.2
3:00 272.6 F 29.1
4:00 272.2 F 28.6

Min 271.4 F 28.2
Max 272.7 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.7 F 0.0
5:00 233.1 F 0.0
10:00 231.4 F 0.0
15:00 221.6 F 0.0
20:00 221.5 F 0.0
25:00 227.1 F 0.0
30:00 227.7 F 0.0
Drying Complete

FILLING: 2:08
HEATING: 18:26
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 00:57:49
CYCLE COMPLETE

04/12/2023 01:53 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____