

-----  
Midmark M11 - v1.0.5  
Total Cycles: 4278

Name: RITTER2  
Sterilizer ID: V2313228  
01/22/2025 11:24 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 128.8 F 0.0  
2:00 128.1 F 1.0  
4:00 131.0 F 3.2  
6:00 159.4 F 6.6  
8:00 181.2 F 10.6  
10:00 197.3 F 4.9  
12:00 213.5 F 7.0  
14:00 225.3 F 11.1  
16:00 229.6 F 6.9  
18:00 238.4 F 11.8  
20:00 237.2 F 9.9  
22:00 249.8 F 15.8  
24:00 260.4 F 21.7  
26:00 270.1 F 28.2

STERILIZING  
mm:ss Degrees PSI  
0:00 272.1 F 29.2  
1:00 272.1 F 29.9  
2:00 271.5 F 29.1  
3:00 272.3 F 29.9  
4:00 271.5 F 29.2

Min 271.3 F 29.1  
Max 272.6 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 198.0 F 0.0  
5:00 179.3 F 0.0  
10:00 181.5 F 0.0  
15:00 178.2 F 0.0  
20:00 175.2 F 0.0  
25:00 183.9 F 0.0  
30:00 199.0 F 0.0  
Drying Complete

FILLING: 0:57  
HEATING: 26:47  
STERILIZING: 4:00  
VENTING: 3:44  
DRYING: 30:00  
TOTAL CYCLE: 01:05:28  
CYCLE COMPLETE

01/22/2025 12:29 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_