
Midmark M11 - v1.0.0
Total Cycles: 4372

Name: RITTER 1
Sterilizer ID: V2178198
03/08/2024 02:14 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 121.8 F 0.0
2:00 123.7 F 1.4
4:00 135.3 F 3.9
6:00 165.7 F 7.0
8:00 194.2 F 10.8
10:00 209.7 F 3.6
12:00 216.2 F 7.0
14:00 226.8 F 11.3
16:00 228.5 F 6.8
18:00 241.0 F 12.4
20:00 239.8 F 10.4
22:00 252.0 F 16.5
24:00 262.5 F 22.6
26:00 272.5 F 29.2

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.4
1:00 272.7 F 29.2
2:00 272.1 F 28.9
3:00 272.8 F 29.4
4:00 271.5 F 28.8

Min 271.0 F 28.3
Max 273.0 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 198.0 F 0.0
5:00 209.3 F 0.0
10:00 220.2 F 0.0
15:00 221.8 F 0.0
20:00 218.9 F 0.0
25:00 216.6 F 0.0
30:00 216.3 F 0.0
Drying Complete

FILLING: 0:60
HEATING: 26:29
STERILIZING: 4:00
VENTING: 3:42
DRYING: 30:00
TOTAL CYCLE: 01:05:11
CYCLE COMPLETE

03/08/2024 03:19 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____