
Midmark M11 - v1.0.0
Total Cycles: 4311

Name: RITTER 1
Sterilizer ID: V2178198
02/16/2024 03:40 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 95.5 F 0.0
2:00 98.3 F 0.8
4:00 109.6 F 2.5
6:00 135.8 F 4.9
8:00 180.6 F 7.8
10:00 190.2 F 11.6
12:00 213.0 F 3.0
14:00 216.2 F 6.6
16:00 225.1 F 10.4
18:00 232.0 F 8.4
20:00 238.7 F 10.8
22:00 235.6 F 8.2
24:00 246.4 F 13.9
26:00 256.7 F 19.4
28:00 266.9 F 25.4

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.6
1:00 272.6 F 29.2
2:00 271.2 F 28.5
3:00 272.4 F 29.3
4:00 271.9 F 28.8

Min 271.1 F 28.5
Max 272.7 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 200.4 F 0.0
5:00 207.7 F 0.0
10:00 216.0 F 0.0
15:00 197.1 F 0.0
20:00 187.4 F 0.0
25:00 182.7 F 0.0
30:00 182.1 F 0.0
Drying Complete

FILLING: 1:02
HEATING: 29:40
STERILIZING: 4:00
VENTING: 3:45
DRYING: 30:00
TOTAL CYCLE: 01:08:28
CYCLE COMPLETE

02/16/2024 04:48 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____