
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 002387
11:50 16/12/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:07
PRESSURIZING 2:13
STERILIZING 3:09
135.9 C 220kPa 3:09
136.2 C 224kPa 3:39
136.3 C 224kPa 4:09
136.3 C 224kPa 4:39
136.4 C 226kPa 5:09
136.4 C 225kPa 5:39
136.3 C 225kPa 6:09
136.2 C 224kPa 6:39
136.2 C 224kPa 6:39
VENTING 6:39
AIR DRYING 7:36
CYCLE COMPLETE 7:50

Digital Signature #
8E0D46808B42F14A
