

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2227

Name: RITTER 1  
Sterilizer ID: V2178198  
08/16/2022 09:43 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 181.9 F 0.0  
2:00 179.3 F 3.2  
4:00 182.5 F 7.8  
6:00 202.1 F 13.0  
8:00 223.6 F 6.6  
10:00 233.3 F 11.5  
12:00 232.7 F 7.8  
14:00 244.7 F 12.1  
16:00 244.2 F 12.2  
18:00 256.4 F 18.6  
20:00 267.3 F 25.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.2  
1:00 272.6 F 28.9  
2:00 271.8 F 28.3  
3:00 271.7 F 28.5  
4:00 272.6 F 28.9  
  
Min 271.3 F 28.0  
Max 272.7 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 203.7 F 0.0  
5:00 216.1 F 0.0  
10:00 232.0 F 0.0  
15:00 229.7 F 0.0  
20:00 222.1 F 0.0  
25:00 222.9 F 0.0  
30:00 234.3 F 0.0  
Drying Complete

FILLING: 1:32  
HEATING: 21:27  
STERILIZING: 4:00  
VENTING: 3:13  
DRYING: 30:00  
TOTAL CYCLE: 01:00:13  
CYCLE COMPLETE

08/16/2022 10:43 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_