
Midmark M11 - v1.0.5
Total Cycles: 3677

Name: RITTER2
Sterilizer ID: V2313228
06/07/2024 11:12 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 166.7 F 0.0
2:00 165.1 F 2.5
4:00 167.8 F 5.6
6:00 186.8 F 9.5
8:00 207.4 F 11.7
10:00 215.9 F 7.1
12:00 227.1 F 11.5
14:00 230.1 F 8.1
16:00 242.4 F 13.1
18:00 242.4 F 12.5
20:00 254.2 F 18.4
22:00 264.5 F 24.8
24:00 271.4 F 29.3

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.4
1:00 272.5 F 30.3
2:00 272.1 F 29.8
3:00 271.1 F 29.3
4:00 271.5 F 30.0

Min 271.0 F 29.3
Max 272.6 F 30.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 201.8 F 0.0
5:00 203.2 F 0.0
10:00 206.3 F 0.0
15:00 199.2 F 0.0
20:00 194.0 F 0.0
25:00 189.5 F 0.0
30:00 186.2 F 0.0
Drying Complete

FILLING: 1:05
HEATING: 24:06
STERILIZING: 4:00
VENTING: 3:45
DRYING: 30:00
TOTAL CYCLE: 01:02:57
CYCLE COMPLETE

06/07/2024 12:14 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____