
Midmark M11 - v1.0.0
Total Cycles: 1666

Name: RITTER 1
Sterilizer ID: V2178198
02/24/2022 07:53 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 75.0 F 0.0
2:00 84.9 F 0.0
4:00 92.9 F 0.3
6:00 122.8 F 1.2
8:00 151.3 F 2.5
10:00 174.9 F 4.4
12:00 192.1 F 7.0
14:00 208.5 F 10.5
16:00 220.7 F 5.2
18:00 231.1 F 8.3
20:00 240.5 F 12.9
22:00 238.1 F 9.7
24:00 238.1 F 8.8
26:00 248.8 F 14.2
28:00 260.7 F 20.7
30:00 271.3 F 27.7

STERILIZING
mm:ss Degrees PSI
0:00 272.0 F 27.8
1:00 272.7 F 28.6
2:00 271.6 F 27.7
3:00 272.7 F 28.5
4:00 271.5 F 27.7

Min 271.5 F 27.7
Max 272.9 F 28.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 175.7 F 0.0
5:00 233.0 F 0.0
10:00 229.3 F 0.0
15:00 232.5 F 0.0
20:00 217.2 F 0.0
25:00 206.6 F 0.0
30:00 204.9 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 30:30
STERILIZING: 4:00
VENTING: 3:04
DRYING: 30:00
TOTAL CYCLE: 01:08:28
CYCLE COMPLETE

02/24/2022 10:20 AM
ACCEPTED: X
REJECTED:
Operator: U02