
Midmark M11 - v1.0.0
Total Cycles: 1786

Name: RITTER 1
Sterilizer ID: V2178198
04/04/2022 07:13 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 71.5 F 0.0
2:00 72.0 F 0.0
4:00 82.3 F 0.0
6:00 110.4 F 0.0
8:00 146.1 F 2.2
10:00 172.5 F 5.4
12:00 190.0 F 8.7
14:00 205.9 F 12.0
16:00 218.5 F 4.5
18:00 227.4 F 8.3
20:00 236.7 F 12.3
22:00 234.6 F 8.5
24:00 244.2 F 11.9
26:00 244.0 F 12.1
28:00 255.7 F 18.3
30:00 266.1 F 24.8

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.2
1:00 272.5 F 28.8
2:00 271.6 F 28.5
3:00 272.3 F 28.6
4:00 271.9 F 28.7

Min 271.3 F 28.1
Max 272.8 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 181.6 F 0.0
5:00 189.4 F 0.0
10:00 181.4 F 0.0
15:00 182.6 F 0.0
20:00 186.8 F 0.0
25:00 186.0 F 0.0
30:00 181.5 F 0.0
Drying Complete

FILLING: 0:44
HEATING: 31:38
STERILIZING: 4:00
VENTING: 3:06
DRYING: 30:00
TOTAL CYCLE: 01:09:28
CYCLE COMPLETE

04/04/2022 08:30 AM
ACCEPTED: X
REJECTED:
Operator: U02