
Midmark M11 - v1.0.0
Total Cycles: 3562

Name: RITTER 1
Sterilizer ID: V2178198
07/21/2023 01:52 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 98.2 F 0.0
2:00 101.8 F 0.0
4:00 114.2 F 0.8
6:00 156.5 F 2.4
8:00 188.6 F 4.7
10:00 198.2 F 8.0
12:00 211.9 F 11.4
14:00 218.5 F 4.7
16:00 226.6 F 8.7
18:00 236.7 F 13.1
20:00 237.0 F 9.8
22:00 238.2 F 9.6
24:00 246.0 F 13.3
26:00 256.9 F 19.0
28:00 266.7 F 25.1
30:00 272.7 F 29.3

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.7
1:00 272.3 F 28.9
2:00 272.6 F 29.3
3:00 271.5 F 28.7
4:00 272.3 F 29.1

Min 271.4 F 28.6
Max 272.8 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.5 F 0.0
5:00 202.7 F 0.0
10:00 217.1 F 0.0
15:00 219.9 F 0.0
20:00 225.1 F 0.0
25:00 231.6 F 0.0
30:00 223.1 F 0.0
Drying Complete

FILLING: 0:52
HEATING: 30:22
STERILIZING: 4:00
VENTING: 3:35
DRYING: 30:00
TOTAL CYCLE: 01:08:50
CYCLE COMPLETE

07/21/2023 03:00 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____