

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1682

Name: RITTER2  
Sterilizer ID: V2313228  
11/03/2022 03:16 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 200.0 F 0.0  
2:00 192.9 F 3.8  
4:00 207.7 F 9.5  
6:00 222.5 F 7.1  
8:00 235.4 F 10.9  
10:00 235.2 F 9.0  
12:00 238.9 F 9.8  
14:00 249.6 F 15.2  
16:00 261.7 F 22.2  
18:00 272.7 F 29.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.7  
1:00 271.9 F 29.0  
2:00 272.0 F 29.4  
3:00 271.5 F 28.7  
4:00 272.1 F 29.5  
  
Min 271.1 F 28.6  
Max 272.5 F 29.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 197.2 F 0.0  
5:00 226.3 F 0.0  
10:00 239.7 F 0.0  
15:00 229.4 F 0.0  
20:00 228.1 F 0.0  
25:00 227.8 F 0.0  
30:00 225.2 F 0.0  
Drying Complete

FILLING: 2:36  
HEATING: 18:22  
STERILIZING: 4:00  
VENTING: 3:11  
DRYING: 30:00  
TOTAL CYCLE: 00:58:09  
CYCLE COMPLETE

11/03/2022 04:14 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_