
Midmark M11 - v1.0.0
Total Cycles: 1145

Name: RITTER 1
Sterilizer ID: V2178198
09/08/2021 06:47 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 71.0 F 0.0
2:00 73.6 F 0.0
4:00 92.3 F 0.0
6:00 109.4 F 0.1
8:00 142.9 F 0.2
10:00 174.5 F 0.3
12:00 196.9 F 2.2
14:00 214.3 F 5.5
16:00 221.7 F 9.0
18:00 232.1 F 12.8
20:00 232.9 F 8.6
22:00 243.0 F 12.2
24:00 242.2 F 11.8
26:00 240.0 F 10.4
28:00 253.5 F 17.2
30:00 265.0 F 23.9

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.3
1:00 272.2 F 28.6
2:00 272.4 F 29.1
3:00 271.6 F 28.2
4:00 272.4 F 28.9

Min 271.5 F 28.2
Max 272.6 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.4 F 0.0
5:00 214.0 F 0.0
10:00 220.8 F 0.0
15:00 215.4 F 0.0
20:00 209.9 F 0.0
25:00 211.2 F 0.0
30:00 209.7 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 31:48
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 01:09:45
CYCLE COMPLETE

09/08/2021 08:02 AM
ACCEPTED: X
REJECTED:
Operator: U01