

-----  
Midmark M11 - v1.0.0  
Total Cycles: 5485

Name: RITTER 1  
Sterilizer ID: V2178198  
04/07/2025 03:36 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 98.5 F 0.0  
2:00 100.3 F 0.8  
4:00 119.1 F 3.0  
6:00 150.5 F 5.7  
8:00 172.2 F 8.8  
10:00 190.6 F 12.0  
12:00 215.4 F 4.3  
14:00 218.3 F 8.2  
16:00 230.5 F 12.1  
18:00 230.7 F 7.4  
20:00 240.6 F 12.3  
22:00 238.3 F 9.7  
24:00 249.8 F 15.2  
26:00 260.0 F 20.9  
28:00 269.5 F 27.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 29.0  
1:00 272.7 F 29.3  
2:00 271.6 F 29.3  
3:00 272.3 F 29.2  
4:00 271.8 F 29.4

Min 271.0 F 28.9  
Max 272.8 F 29.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 174.7 F 0.0  
5:00 233.7 F 0.0  
10:00 227.4 F 0.0  
15:00 232.4 F 0.0  
20:00 223.0 F 0.0  
25:00 212.5 F 0.0  
30:00 190.6 F 0.0  
Drying Complete

FILLING: 0:48  
HEATING: 29:52  
STERILIZING: 4:00  
VENTING: 3:53  
DRYING: 30:00  
TOTAL CYCLE: 01:08:34  
CYCLE COMPLETE

04/07/2025 04:44 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_