
Midmark M11 - v1.0.5
Total Cycles: 959

Name: RITTER2
Sterilizer ID: V2313228
04/06/2022 06:44 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 73.6 F 0.0
2:00 75.8 F 0.0
4:00 83.9 F 0.0
6:00 101.6 F 1.6
8:00 143.9 F 4.8
10:00 176.0 F 8.1
12:00 191.7 F 12.0
14:00 211.7 F 5.0
16:00 228.3 F 8.7
18:00 237.9 F 12.9
20:00 233.9 F 9.2
22:00 241.5 F 10.7
24:00 244.3 F 12.9
26:00 255.8 F 19.1
28:00 266.2 F 25.5

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.2
1:00 272.7 F 29.9
2:00 271.5 F 29.1
3:00 272.0 F 29.9
4:00 272.6 F 29.6

Min 271.1 F 29.0
Max 272.9 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.9 F 0.0
5:00 193.2 F 0.0
10:00 198.2 F 0.0
15:00 210.4 F 0.0
20:00 220.8 F 0.0
25:00 225.8 F 0.0
30:00 220.0 F 0.0
Drying Complete

FILLING: 0:48
HEATING: 29:44
STERILIZING: 4:00
VENTING: 3:05
DRYING: 30:00
TOTAL CYCLE: 01:07:37
CYCLE COMPLETE

04/06/2022 07:51 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____