
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 002026
12:55 12/09/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:50
PRESSURIZING 1:58
STERILIZING 2:59
136.4 C 222kPa 2:59
136.4 C 226kPa 3:29
136.1 C 224kPa 3:59
136.3 C 224kPa 4:29
136.3 C 224kPa 4:59
136.3 C 225kPa 5:29
136.3 C 224kPa 5:59
136.3 C 225kPa 6:29
136.3 C 225kPa 6:29
VENTING 6:29
AIR DRYING 7:21
CYCLE COMPLETE 37:21

Digital Signature #
990253C9A6071AA3
