
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000933
13:08 08/12/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:48
PRESSURIZING 3:49
STERILIZING 4:56
136.4 C 225kPa 4:56
136.1 C 222kPa 5:26
136.4 C 224kPa 5:56
136.4 C 223kPa 6:26
136.4 C 224kPa 6:56
136.3 C 223kPa 7:26
136.2 C 221kPa 7:56
136.3 C 221kPa 8:26
136.3 C 221kPa 8:26
VENTING 8:26
AIR DRYING 9:21
CYCLE COMPLETE 9:22

Digital Signature #
E7516E7354905EC0
