
Midmark M11 - v1.0.5
Total Cycles: 1605

Name: RITTER2
Sterilizer ID: V2313228
10/17/2022 09:49 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 203.7 F 0.0
2:00 199.5 F 3.7
4:00 199.8 F 9.0
6:00 218.6 F 8.2
8:00 229.5 F 10.1
10:00 233.4 F 8.4
12:00 242.7 F 12.2
14:00 244.6 F 13.2
16:00 257.3 F 19.9
18:00 268.5 F 27.0

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.2
1:00 272.4 F 29.8
2:00 271.5 F 29.0
3:00 272.5 F 29.9
4:00 272.1 F 29.4

Min 271.4 F 29.0
Max 272.7 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.4 F 0.0
5:00 216.3 F 0.0
10:00 233.0 F 0.0
15:00 232.1 F 0.0
20:00 224.8 F 0.0
25:00 221.2 F 0.0
30:00 219.3 F 0.0
Drying Complete

FILLING: 2:09
HEATING: 19:13
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:58:33
CYCLE COMPLETE

10/17/2022 10:47 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____