
Midmark M11 - v1.0.5
Total Cycles: 1879

Name: RITTER2
Sterilizer ID: V2313228
12/22/2022 10:12 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 189.5 F 0.0
2:00 189.7 F 3.4
4:00 193.7 F 8.6
6:00 217.0 F 8.0
8:00 228.7 F 10.0
10:00 234.3 F 8.1
12:00 242.9 F 11.6
14:00 245.3 F 13.6
16:00 258.5 F 20.7
18:00 270.1 F 28.3

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.3
1:00 272.8 F 30.0
2:00 271.9 F 29.3
3:00 272.1 F 29.9
4:00 272.6 F 29.9

Min 271.5 F 29.1
Max 272.8 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.0 F 0.0
5:00 213.6 F 0.0
10:00 229.2 F 0.0
15:00 233.1 F 0.0
20:00 226.0 F 0.0
25:00 225.7 F 0.0
30:00 210.5 F 0.0
Drying Complete

FILLING: 2:10
HEATING: 18:54
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 00:58:14
CYCLE COMPLETE

12/22/2022 11:10 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____