

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4499

Name: RITTER 1  
Sterilizer ID: V2178198  
04/26/2024 04:04 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 180.9 F 0.0  
2:00 189.6 F 4.9  
4:00 211.2 F 10.7  
6:00 225.7 F 6.1  
8:00 240.5 F 12.0  
10:00 238.6 F 9.7  
12:00 236.4 F 8.1  
14:00 251.9 F 15.7  
16:00 264.6 F 23.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 27.8  
1:00 272.1 F 28.0  
2:00 272.7 F 28.6  
3:00 271.7 F 27.9  
4:00 272.5 F 28.2  
  
Min 271.6 F 27.7  
Max 272.8 F 28.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 201.2 F 0.0  
5:00 229.2 F 0.0  
10:00 223.6 F 0.0  
15:00 218.6 F 0.0  
20:00 212.4 F 0.0  
25:00 226.8 F 0.0  
30:00 215.6 F 0.0  
Drying Complete

FILLING: 2:37  
HEATING: 17:45  
STERILIZING: 4:00  
VENTING: 3:07  
DRYING: 30:00  
TOTAL CYCLE: 00:57:28  
CYCLE COMPLETE

04/26/2024 05:01 PM  
ACCEPTED:  
REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_