
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 006839
9:06 19/01/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:03
PRESSURIZING 3:05
STERILIZING 4:12
136.4 C 222kPa 4:12
136.2 C 221kPa 4:42
136.3 C 222kPa 5:12
135.9 C 219kPa 5:42
135.9 C 221kPa 6:12
136.4 C 222kPa 6:42
136.4 C 220kPa 7:12
136.4 C 222kPa 7:42
136.4 C 222kPa 7:42
VENTING 7:43
AIR DRYING 8:31
CYCLE COMPLETE 38:31

Digital Signature #
4FBE14745BCFE3D4
