
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 006365
14:21 26/09/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 0:54
PRESSURIZING 1:00
STERILIZING 1:58
136.4 C 223kPa 1:58
136.4 C 224kPa 2:28
136.4 C 225kPa 2:58
136.4 C 225kPa 3:28
136.4 C 223kPa 3:58
136.4 C 223kPa 4:28
136.4 C 222kPa 4:58
136.4 C 222kPa 5:28
136.4 C 222kPa 5:28
VENTING 5:28
AIR DRYING 6:20
CYCLE COMPLETE 6:25

Digital Signature #
47538FE14AFC48EC
