
Midmark M11 - v1.0.0
Total Cycles: 1712

Name: RITTER 1
Sterilizer ID: V2178198
03/14/2022 10:42 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 180.5 F 0.0
2:00 183.4 F 2.9
4:00 190.2 F 7.3
6:00 214.0 F 12.7
8:00 224.6 F 6.9
10:00 237.5 F 12.1
12:00 236.5 F 9.3
14:00 239.7 F 9.2
16:00 248.5 F 14.3
18:00 260.6 F 21.1
20:00 271.6 F 28.3

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.2
1:00 272.6 F 28.9
2:00 271.6 F 28.1
3:00 272.4 F 29.0
4:00 272.3 F 28.7

Min 271.4 F 28.1
Max 272.7 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.0 F 0.0
5:00 213.4 F 0.0
10:00 224.8 F 0.0
15:00 235.1 F 0.0
20:00 230.6 F 0.0
25:00 229.1 F 0.0
30:00 229.0 F 0.0
Drying Complete

FILLING: 1:12
HEATING: 20:38
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 00:58:57
CYCLE COMPLETE

03/14/2022 11:40 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____