

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3839

Name: RITTER 1  
Sterilizer ID: V2178198  
09/29/2023 09:21 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 164.9 F 0.0  
2:00 162.3 F 0.0  
4:00 171.3 F 1.0  
6:00 197.8 F 4.3  
8:00 208.4 F 8.7  
10:00 223.2 F 13.4  
12:00 229.0 F 8.8  
14:00 237.8 F 11.5  
16:00 241.3 F 11.6  
18:00 239.4 F 10.1  
20:00 252.7 F 16.6  
22:00 264.0 F 23.2  
24:00 271.6 F 28.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.1  
1:00 272.5 F 28.9  
2:00 271.9 F 28.3  
3:00 271.5 F 28.4  
4:00 272.4 F 28.9

Min 271.2 F 28.0  
Max 272.6 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 196.3 F 0.0  
5:00 208.1 F 0.0  
10:00 208.8 F 0.0  
15:00 206.2 F 0.0  
20:00 226.4 F 0.0  
25:00 230.1 F 0.0  
30:00 230.8 F 0.0  
Drying Complete

FILLING: 0:50  
HEATING: 24:08  
STERILIZING: 4:00  
VENTING: 3:27  
DRYING: 30:00  
TOTAL CYCLE: 01:02:25  
CYCLE COMPLETE

09/29/2023 10:23 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_