
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 002494
8:43 10/02/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:13
PRESSURIZING 3:16
STERILIZING 4:20
136.4 C 223kPa 4:20
136.1 C 221kPa 4:50
136.3 C 221kPa 5:20
136.3 C 223kPa 5:50
136.3 C 223kPa 6:20
136.4 C 225kPa 6:50
136.4 C 223kPa 7:20
136.3 C 223kPa 7:50
136.3 C 223kPa 7:50
VENTING 7:50
AIR DRYING 8:41
CYCLE COMPLETE 24:48

Digital Signature #
C2F3048CCB02ED44
