
Midmark M11 - v1.0.5
Total Cycles: 4151

Name: RITTER2
Sterilizer ID: V2313228
11/21/2024 05:42 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 153.9 F 0.0
2:00 156.1 F 1.7
4:00 179.3 F 5.7
6:00 205.4 F 10.1
8:00 219.1 F 5.9
10:00 229.9 F 10.6
12:00 230.5 F 7.7
14:00 239.4 F 11.5
16:00 244.3 F 13.4
18:00 257.4 F 20.2
20:00 268.4 F 27.4

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 29.4
1:00 272.2 F 29.8
2:00 271.5 F 29.8
3:00 271.8 F 29.6
4:00 272.1 F 30.3

Min 271.0 F 29.2
Max 273.1 F 30.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 197.4 F 0.0
5:00 230.2 F 0.0
10:00 218.5 F 0.0
15:00 229.8 F 0.0
20:00 224.7 F 0.0
25:00 228.9 F 0.0
30:00 226.7 F 0.0
Drying Complete

FILLING: 1:13
HEATING: 21:07
STERILIZING: 4:00
VENTING: 3:32
DRYING: 30:00
TOTAL CYCLE: 00:59:52
CYCLE COMPLETE

11/21/2024 06:41 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____