

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2589

Name: RITTER2  
Sterilizer ID: V2313228  
06/29/2023 01:29 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 155.1 F 0.0  
2:00 156.6 F 1.6  
4:00 181.8 F 5.8  
6:00 203.2 F 10.8  
8:00 217.6 F 6.1  
10:00 231.8 F 11.7  
12:00 233.4 F 9.1  
14:00 238.0 F 9.7  
16:00 247.5 F 14.6  
18:00 260.1 F 21.6  
20:00 271.1 F 28.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 29.2  
1:00 272.7 F 30.0  
2:00 272.1 F 29.4  
3:00 271.4 F 29.3  
4:00 272.6 F 30.0

Min 271.3 F 29.0  
Max 272.9 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 205.2 F 0.0  
5:00 207.1 F 0.0  
10:00 212.4 F 0.0  
15:00 219.8 F 0.0  
20:00 220.4 F 0.0  
25:00 220.6 F 0.0  
30:00 222.7 F 0.0  
Drying Complete

FILLING: 1:26  
HEATING: 20:44  
STERILIZING: 4:00  
VENTING: 3:16  
DRYING: 30:00  
TOTAL CYCLE: 00:59:26  
CYCLE COMPLETE

06/29/2023 02:28 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_